



## QUALITY CONTROL CERTIFICATE

**Name of Product:** Multi-Drug One Step Cup (Urine)

**Lot Number:** DOAA2309257

**Expiration Date:** 2025-09-25

**Lot Size:** 124200 PCS

### Product Test Results

ITEM	TEST	STANDARDS	RESULT
BZO300ng/ml	0 ng/ml	Negative (-)	PASS
	-50%	Negative (-)	PASS
	+50%	Positive (+)	PASS
COC300ng/ml	0 ng/ml	Negative (-)	PASS
	-50%	Negative (-)	PASS
	+50%	Positive (+)	PASS
THC50ng/ml	0ng/ml	Negative(-)	PASS
	-50%	Negative(-)	PASS
	+50%	Positive(+)	PASS

No radioactive material of any kind is utilized in the product, or in the manufacturing of the product.

The product specified above does not contain any virus, reagent by-product of the same, or metabolic by-product of Hepatitis A, B, C, D.

These testing procedures adhere to strict quality control and have a low probability of error.

**Reviewed by: QC Supervisor**

**Signature:**

**Date:** 2023-10-23

BTNX Inc.  
www.btnx.com  
Tel. 1.888.339.9964  
Fax 905.944.0406